

# INDIAN INSTITUTE OF TECHNOLOGY GOA

At Goa Engineering College Campus

Farmagudi, Ponda, Goa 403401

E-mail: [purchase@iitgoa.ac.in](mailto:purchase@iitgoa.ac.in)

**Enquiry No: IITGOA/2018-19/056**

**Date: 01/01/2019**

IIT Goa invites sealed quotations in two bid format for the supply of below mentioned item.

Sl. No.	Description of Item	Qty
1	Flow/Pressure Calibrator (Detailed Specifications Attached)	01

## Terms and conditions:

1. Quotation must be valid for at least 60 days.
2. The GSTIN should invariably be mentioned in your offer.
3. Kindly attach a compliance certificate along with the technical quote.
4. Supplier should provide at least one-year warranty for the above-mentioned item.
5. The supplier should provide a list of customers of previous supply of a similar/ same range of equipment to IIT's / NIT's / Universities with contact details. Copies of orders received from the reputed firms on bidding firm need to be submitted.
6. Prices should be quoted in Indian Rupees inclusive of taxes and any shipping charges.
7. Delivery and installation should be made within 4 weeks of getting a confirmed order.
8. Payment: Within 30 days after the delivery and installation of the item at IIT GOA.
9. The suppliers shall provide the banking details along with their quote on their letterhead duly signed and stamped.
10. IIT Goa reserves the right to accept and/or reject any/all bids without assigning any reason.
11. Quotations shall be submitted in two parts;
  - 1) **Part – I (Technical)** should contain all the technical details and specification of the product. It should contain unpriced bid along with terms and conditions, compliance certificates, proprietary certificates (if applicable), any other certificates/details etc. This envelope should be marked as “Technical Bid”
  - 2) **Part -II (Financial)** The financial bid of the above item should be in a sealed envelope marked as “Financial Bid” and should contain financial terms and conditions.
12. For any clarification, you may kindly contact Dr. Anirudha Ambekar (e-mail: [anirudha@iitgoa.ac.in](mailto:anirudha@iitgoa.ac.in)) and Stores & Purchase Department (email: [purchase@iitgoa.ac.in](mailto:purchase@iitgoa.ac.in)) till 11/01/2019.

13. All sealed quotations must reach to the Assistant Registrar (Stores & Purchase), IIT Goa, at Goa College of Engineering Campus, Farmagudi, Ponda, Goa by 17.00 Hrs on or before 22<sup>nd</sup> January, 2019”.

Sd/-

Asst. Registrar (S&P)

### **Specifications of Pressure /Flow calibrator**

- Range: 0 -200 mm of water column
- Accuracy: 0.1% of reading
- Resolution: 1:200000
- Static pressure: -1 bar to 4 bar
- Storage temperature: -10 deg C to 70 deg C
- Pneumatic ports: 4mm O/D x 3mm I/D push-on tube
- Interface: USB Type A: memory stick
- USB Type B: PC connection
- Display: Blue backlit graphic display
- Internal storage: 30000 readings
- Max operating temperature: 5 to 45 deg C
- With Company Traceable Certificate (Working Standard of Company)