

Enquiry No. IITGOA/2018-19/038

Date: 27/12/2018

Corrigendum to the Tender for supply of Communication Laboratory Items vide Enquiry No. IITGOA/2018-19/038 dtd. 11/12/2018.

For the tender for supply of Communication Laboratory Items at IIT Goa, the following clauses / paragraphs have been modified;

1) DIGITAL STORAGE OSCILLOSCOPE

TECHNICAL SPECIFICATIONS		
1	Bandwidth	70MHz
2	Max Sample Rate	1 GSa/s (Should be available for each channel simultaneously)
3	Channels	2 Channels/4 Channels Quotation must be given separately for each.
4	USB Connector	
5	Memory Card	
6	Standard 5-year warranty	
7	Should have LAN/Wireless Connectivity	

2) AFG

TECHNICAL SPECIFICATIONS		
1	Dual-channel, 25MHz bandwidth with 1mVpp to 10Vpp output; 14-bit vertical resolution and 1uHz frequency resolution	
2	50 built-in arbitrary waveforms with continuous, modulation, sweep and burst modes	
3	Dedicated "3.5 to 3.95" TFT LCD display, easy-to-navigate front panel and menu	
4	Built-in 200MHz, 6-digit frequency counter	
5	Standard 5-year warranty	

All other terms will remain same.

Sd/-
Assistant Registrar