

INDIAN INSTITUTE OF TECHNOLOGY GOA
Goa College of Engineering Campus, Farmagudi, Ponda -403401, Goa

Enquiry No. IITGOA/2020-21/013

Date: 16/11/2020

**Corrigendum to the tender for Supply of Kugelrohr Apparatus vide Enquiry No.
IITGOA/2020-21/013 dtd. 28/10/2020.**

For the tender of Supply of Kugelrohr Apparatus at IIT Goa, the following clauses / paragraphs have been modified in the technical specification:

S. No.	Old Specification	New Specification
11.	Control unit with large 4.3 inches LCD display for centrally controlling all process parameters.	Control unit should be with display for centrally controlling vacuum pump
14.	Automatic aeration when pressure is above 1400 mbar.	S. No. 14 can be ignored
15.	Should have integrated solvent database for setting up dynamic distillation conditions.	S. No. 15 can be ignored
24.	Should be speed-controlled pump and stops after reaching desired vacuum pressure.	Should be Automatic On/Off-function pump and stops after reaching desired vacuum Pressure

All other terms will remain same.

Sd/-
Assistant Registrar