

**INDIAN INSTITUTE OF TECHNOLOGY GOA**  
Goa College of Engineering Campus, Farmagudi, Ponda -403401, Goa

---

Enquiry No. IITGOA/2018-19/074

Date: 24/01/2019

**Corrigendum to the Tender for supply of Atomic Force Microscope vide Enquiry No. IITGOA/2018-19/074 dtd. 16/01/2019.**

For the tender for supply of Atomic Force Microscope at IIT Goa, the following clauses / paragraphs have been modified:

# In Annexure – B, under scanners section, point no 4.1 a) should be read as **“System must automatically scan the sample in XY and the tip automatically in Z with appropriate computer control”** instead of “System must scan the sample in XY and the tip in Z”.

All other terms will remain same.

Sd/-  
Assistant Registrar