

Enquiry No. IITGOA/2018-19/093

Date: 06/02/2019

**Corrigendum to the Tender for supply of Notch Filter set for Raman head of Spectrometer
vide Enquiry No. IITGOA/2018-19/093 dtd. 05/02/2019.**

For the tender for supply of Notch Filter set for Raman head of Spectrometer at IIT Goa, the following clauses / paragraphs have been modified:

In specifications of (1),

- 1) “Angle tunable to compensate small change of ± 0.1 nm in the laser wavelength”
must be read as
“Angle tunable to compensate small change of ± 0.3 nm in the laser wavelength”
- 2) “Input/output angles should be higher than 10 degree”
must be read as
“Input/output angles should not be higher than 10 degree”

In specifications of (2),

- 1) “Input/output angles should be higher than 15 degree”
must be read as
“Input/output angles should not be higher than 15 degree”

All other terms will remain same.

Sd/-
Assistant Registrar